



EXPLOSION PROTECTION CERTIFICATE OF CONFORMITY

Cert NO.GYB19.1462X

This is to certify that the product

Temperature transmitter

manufactured by Microcyber Corporation

(Address:17-8 Wensu Street, Hunnan New District, Shenyang.)

which model is NCS-TT106H, NCS-TT106H-R1

Ex marking Ex ia IIC T4 Ga

product standard Q/BW02-2018

drawing number TT106HBA-19-05-00、T106H-R1BA-19-04-00

has been inspected and certified by NEPSI, and that it conforms
to GB 3836.1-2010,GB 3836.4-2010,GB 3836.20-2010

This Approval shall remain in force until 2024.05.14

Remarks
1.Conditions for safe use are specified in the instructions.
2.Symbol "X" placed after the certification number denotes specific conditions of use,
which are specified in the attachment to this certificate.
3.Intrinsic safety parameters specified in the attachment to this certificate.

Director

National Supervision and Inspection Centre for
Explosion Protection and Safety of Instrumentation

Issued Date 2019.05.15



This Certificate is valid for products compatible with the documents and samples approved by NEPSI.

103 Cao Bao Road
Shanghai 200233, China

<http://www.nepsi.org.cn>
Email: info@nepsi.org.cn

Tel: +86 21 64368180
Fax: +86 21 64844580